

CALIBRATION LABORATORIES

NVLAP LAB CODE 600214-0

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**


<p><b>Additel Corporation</b> 2900 Saturn Street Brea, CA 92821 Eric Chavier Phone: 714-998-6899 Fax: 714-998-6999 E-mail: <a href="mailto:eric.chavier@additel.com">eric.chavier@additel.com</a> URL: <a href="http://additel.com">http://additel.com</a></p>	<p><b>Fields of Calibration</b> Electromagnetics Thermodynamics Time &amp; Frequency</p> <p>This laboratory is compliant to ANSI/NCSL Z540-1-1994; Part 1. (NVLAP Code: 20/A01)</p>
--	---

**CALIBRATION AND MEASUREMENT CAPABILITY (CMC) <sup>Note 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks
<b>ELECTROMAGNETIC - DC/LOW FREQUENCY</b>			
<b>DC RESISTANCE AND CURRENT (20/E05)</b>			
DC Resistance – Generate, Fixed Instrument Based	1 Ω	103 μΩ	Fluke 5730A
	1.9 Ω	180 μΩ	
	10 Ω	250 μΩ	
	19 Ω	475 μΩ	
	100 Ω	1.0 mΩ	
	190 Ω	1.9 mΩ	
	1 kΩ	6.8 mΩ	
	1.9 kΩ	13 mΩ	
	10 kΩ	68 mΩ	
	19 kΩ	130 mΩ	
	100 kΩ	850 mΩ	
	190 kΩ	1.70 Ω	
	1 MΩ	13.3 Ω	
	1.9 MΩ	43.7 Ω	
	10 MΩ	420 Ω	
	19 MΩ	917 Ω	
100 MΩ	13.2 kΩ		

2023-12-19 through 2024-12-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

**CALIBRATION LABORATORIES**


**NVLAP LAB CODE 600214-0**

**CALIBRATION AND MEASUREMENT CAPABILITY (CMC) <sup>Note 1,2</sup>**

<b>Measured Parameter or Device Calibrated</b>	<b>Range</b>	<b>Expanded Uncertainty <sup>Note 3,5</sup></b>	<b>Remarks</b>
DC Resistance – Generate, Fixed Resistors	1 Ω	1.5 μΩ/Ω	Additel ADT280-PRS Reference Resistors with Additel ADT286 Ratio mode
	25 Ω	1.4 μΩ/Ω	
	50 Ω	1.5 μΩ/Ω	
	100 Ω	1.2 μΩ/Ω	
	200 Ω	1.5 μΩ/Ω	
	400 Ω	1.8 μΩ/Ω	
	1 kΩ	1.2 μΩ/Ω	
	2 kΩ	1.5 μΩ/Ω	
	4 kΩ	1.8 μΩ/Ω	
	10 kΩ	1.2 μΩ/Ω	
DC Resistance – Generate, Variable Instrument Based	10 Ω, 400-Ω Range	12 mΩ	Fluke 7526A
	100 Ω, 400-Ω Range	12 mΩ	
	400 Ω, 400-Ω Range	10 mΩ	
	10 Ω, 4k Ω Range	230 mΩ	
	100 Ω, 4k Ω Range	230 mΩ	
	400 Ω, 4k Ω Range	231 mΩ	
DC Resistance - Measure	0 Ω to 10 Ω	11.57 μΩ/Ω + 50 μΩ	Agilent 3458A opt 2
	>10 Ω to 100 Ω	9.2 μΩ/Ω + 500 μΩ	
	>100 Ω to 1 kΩ	8.6 μΩ/Ω + 500 μΩ	
	>1k Ω to 10 kΩ	6.86 μΩ/Ω + 5 mΩ	

2023-12-19 through 2024-12-31

*Effective dates*



For the National Voluntary Laboratory Accreditation Program

**CALIBRATION LABORATORIES**


**NVLAP LAB CODE 600214-0**

**CALIBRATION AND MEASUREMENT CAPABILITY (CMC) <sup>Note 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks
DC Current - Generate	0 mA to 220 $\mu$ A > 220 $\mu$ A to 2.2 mA > 2.2 mA to 22 mA > 22 mA to 220 mA > 220 mA to 2.2 A	40.46 $\mu$ A/A + 6 nA 45.20 $\mu$ A/A + 7 nA 45.20 $\mu$ A/A + 40 nA 44.16 $\mu$ A/A + 0.7 $\mu$ A 93.04 $\mu$ A/A + 12 $\mu$ A	Fluke 5730A
DC Current – Measure	0 to 100 $\mu$ A > 100 $\mu$ A to 1.0 mA > 1.0 mA to 10 mA > 10 mA to 100 mA > 100 mA to 1 A	23.77 $\mu$ A/A + 2.1 nA 21.53 $\mu$ A/A + 6 nA 21.54 $\mu$ A/A + 60 nA 26.33 $\mu$ A/A + 0.6 $\mu$ A 86.75 $\mu$ A/A + 12 $\mu$ A	Agilent 3458A opt 2
<b>DC VOLTAGE (20/E06)</b>			
DC Voltage - Generate	0 mV to 220 mV >220 mV to 2.2 V >2.2 V to 11 V >11 V to 22 V >22V to 220 V >220 to 1100 V	9.19 $\mu$ V/V + 0.4 $\mu$ V 5.25 $\mu$ V/V + 0.7 $\mu$ V 3.47 $\mu$ V/V + 2.5 $\mu$ V 3.47 $\mu$ V/V + 4 $\mu$ V 4.92 $\mu$ V/V + 40 $\mu$ V 6.38 $\mu$ V/V + 400 $\mu$ V	Fluke 5730A
DC Voltage - Measure	0 mV to 100 mV >100 mV to 1 V >1 V to 10 V >10 V to 100 V >100V to 1000 V	7.9 $\mu$ V/V + 0.4 $\mu$ V 3.9 $\mu$ V/V + 0.4 $\mu$ V 3.83 $\mu$ V/V + 0.6 $\mu$ V 24.85 $\mu$ V/V + 40 $\mu$ V 14.63 $\mu$ V/V + 110 $\mu$ V	Agilent 3458A opt 2

2023-12-19 through 2024-12-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 600214-0

CALIBRATION AND MEASUREMENT CAPABILITY (CMC) <sup>Note 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks	
Measured Parameter or Device Calibrated	Range	Frequency	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks
<b>LF AC VOLTAGE (20/E09)</b>				
AC Voltage - Generate	0 mV to 2.2 mV	10 Hz to 40 Hz	767.27 $\mu\text{V/V} + 4.0 \mu\text{V}$	Fluke 5730A
		40 Hz to 20 kHz	616.32 $\mu\text{V/V} + 4.0 \mu\text{V}$	
	>2.2 mV to 22 mV	10 Hz to 40 Hz	312.51 $\mu\text{V/V} + 4.0 \mu\text{V}$	
		40 Hz to 20 kHz	145.96 $\mu\text{V/V} + 4.0 \mu\text{V}$	
	>22 mV to 220 mV	10 Hz to 40 Hz	257.70 $\mu\text{V/V} + 12.0 \mu\text{V}$	
		40 Hz to 20 kHz	72.32 $\mu\text{V/V} + 7.0 \mu\text{V}$	
	>220 mV to 2.2 V	10 Hz to 40 Hz	241.94 $\mu\text{V/V} + 40.0 \mu\text{V}$	
		40 Hz to 20 kHz	46.80 $\mu\text{V/V} + 8.0 \mu\text{V}$	
AC Voltage - Measure	>2.2 V to 22 V	10 Hz to 40 Hz	247.90 $\mu\text{V/V} + 400 \mu\text{V}$	Agilent 3458A opt. 2
		40 Hz to 20 kHz	54.34 $\mu\text{V/V} + 50 \mu\text{V}$	
	>22 V to 220 V	10 Hz to 40 Hz	253.39 $\mu\text{V/V} + 4.0 \text{mV}$	
		40 Hz to 20 kHz	65.83 $\mu\text{V/V} + 0.6 \text{mV}$	
	>220 V to 1100 V	50 Hz to 1 kHz	76.07 $\mu\text{V/V} + 3.5 \text{mV}$	
	0 mV to 10 mV	1 Hz to 40 Hz	421.20 $\mu\text{V/V} + 3.0 \mu\text{V}$	
		40 Hz to 1 kHz	377.58 $\mu\text{V/V} + 1.1 \mu\text{V}$	
		1 kHz to 20 kHz	403.57 $\mu\text{V/V} + 1.1 \mu\text{V}$	
	> 10 mV to 100 mV	1 Hz to 40 Hz	79.71 $\mu\text{V/V} + 4.0 \mu\text{V}$	
		40 Hz to 1 kHz	66.73 $\mu\text{V/V} + 2.0 \mu\text{V}$	
		1 kHz to 20 kHz	104.10 $\mu\text{V/V} + 2.0 \mu\text{V}$	
	> 100 mV to 1 V	1 Hz to 40 Hz	56.81 $\mu\text{V/V} + 40.0 \mu\text{V}$	
		40 Hz to 1 kHz	53.45 $\mu\text{V/V} + 20.0 \mu\text{V}$	
		1 kHz to 20 kHz	105.72 $\mu\text{V/V} + 20.0 \mu\text{V}$	
	> 1 V to 10 V	1 Hz to 40 Hz	72.27 $\mu\text{V/V} + 400.0 \mu\text{V}$	
		40 Hz to 1 kHz	66.05 $\mu\text{V/V} + 200.0 \mu\text{V}$	
	1 kHz to 20 kHz	104.45 $\mu\text{V/V} + 200.0 \mu\text{V}$		
> 10 V to 100 V	1 Hz to 40 Hz	144.63 $\mu\text{V/V} + 4.0 \text{mV}$		
	40 Hz to 1 kHz	144.54 $\mu\text{V/V} + 2.0 \text{mV}$		
	1 kHz to 20 kHz	144.50 $\mu\text{V/V} + 2.0 \text{mV}$		
> 100 V to 1000 V	40 Hz to 1 kHz	294.34 $\mu\text{V/V} + 20.0 \text{mV}$		



2023-12-19 through 2024-12-31

Effective dates

For the National Voluntary Laboratory Accreditation Program

**CALIBRATION LABORATORIES**


**NVLAP LAB CODE 600214-0**

**CALIBRATION AND MEASUREMENT CAPABILITY (CMC) <sup>Note 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks
<b>TIME &amp; FREQUENCY</b>			
<b>FREQUENCY DISSEMINATION (20/F01)</b>			
Frequency - Generate	> 100 Hz to 50 kHz	1 μHz/ Hz + 2 nHz	Keysight 33512B
Frequency - Measure	> 100 Hz to 50 kHz	1 μHz/ Hz + 2 nHz	Keysight 53220A
<b>THERMODYNAMIC</b>			
<b>PRESSURE (20/T05)</b>			
Absolute Pressure Source - Pneumatic	5 kPa to 360 kPa	0.0011 % + 0.64 Pa	DHI PG7601 (10kPa)
	100 kPa to 7200 kPa	0.0017 % + 6.58 Pa	DHI PG7601 (200kPa)
	2 MPa to 72 MPa	0.0037% + 164 Pa	DHI PG7202 (2MPa)
Gauge Pressure Source – Pneumatic <sup>Note 7</sup>	0 Pa to 750 Pa	0.12 Pa	Fluke 7250LP
	750 Pa to 7500 Pa	0.01 %	Fluke 7250LP
	5 kPa to 360 kPa	0.0007 % + 0.53 Pa	DHI PG7601 (10kPa)
	100 kPa to 7200 kPa	0.002 % + 6.7 Pa	DHI PG7601 (200kPa)
	2 MPa to 72 MPa	0.0037%+ 164 Pa	DHI PG7202 (2MPa)
Gauge Pressure Source - Differential	-95 kPa to 10 kPa	0.003 % + 0.79 Pa	DHI PG7601 (10kPa)
Gauge Pressure Source - Hydraulic	7 MPa to 275 MPa	0.0053% + 356 Pa	Minyu KY250
	7 MPa to 20 MPa	3.1 kPa	Fluke P3860-PS
	>20 MPa to 415 MPa	0.017%	Fluke P3860-PSI
<b>RESISTANCE THERMOMETRY (20/T07)</b>			
Drywell Calibrators	-40 °C to 0 °C	0.016 °C	Direct Comparison to SPRT
	>0 °C to 50 °C	0.011 °C	
	>50 °C to 155 °C	0.010 °C	
	>155 °C to 300 °C	0.019 °C	
	>300 °C to 450 °C	0.034 °C	
	>450 °C to 550 °C	0.053 °C	
	>550 °C to 660 °C	0.060 °C	

2023-12-19 through 2024-12-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

**CALIBRATION LABORATORIES**


**NVLAP LAB CODE 600214-0**

**CALIBRATION AND MEASUREMENT CAPABILITY (CMC) <sup>Note 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks
DryWell Calibrators	25 °C to 100 °C >100 °C to 300 °C >300 °C to 600 °C >600 °C to 900 °C >900 °C to 1210 °C	0.40 °C 0.45 °C 0.50 °C 0.55 °C 0.85 °C	Direct Comparison to Reference grade Type S Thermocouple
<b>TEMPERATURE INDICATORS (20/T08)</b>			
RTD Simulation - Measure	-180 °C to 0 °C >0 °C to 800 °C	0.039 °C 0.24 °C	Fluke 7526A PT385 (100 Ω) Measure Mode
RTD Simulation - Generate	-180 °C to 800 °C	0.057 °C	
Thermocouple Simulation – Generate and Measure Type K	-200 °C to 0 °C > 0 °C to 660 °C > 660 °C to 1300 °C	0.12 °C 0.078 °C 0.10 °C	Fluke 7526A Source and Measure
Triple Point	0.01 °C	0.8 mK	Triple Point Water Cell with Accumac SPRT
<b>END</b>			

2023-12-19 through 2024-12-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

**Notes**

**Note 1:** A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

**Note 2:** Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

**Note 3:** The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of  $k = 2$ . However, laboratories may report a coverage factor different than  $k = 2$  to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

**Note 3a:** The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

**Note 3b:** As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

**Note 3c:** As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.5. of NIST Handbook 150, Procedures and General Requirements.

**Note 3d:** CMC expanded uncertainties include repeatability of best existing device (BED).

**Note 4:** Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)


**Note 5:** Uncertainty values listed with percent (%) are percent of reading or generated value unless otherwise noted.

**Note 6:** NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

**Note 7:** Uncertainty applies to positive and negative pressures.

2023-12-19 through 2024-12-31

Effective dates



For the National Voluntary Laboratory Accreditation Program